

4-5-1988

# Final Report on the Construction of Instrumentation to Perform In-House Testing of Torque Wrenches

Erik A. Perry

*Indiana University - Purdue University Fort Wayne*

Follow this and additional works at: [http://opus.ipfw.edu/etcs\\_seniorproj\\_mctid](http://opus.ipfw.edu/etcs_seniorproj_mctid)

---

## Opus Citation

Erik A. Perry (1988). Final Report on the Construction of Instrumentation to Perform In-House Testing of Torque Wrenches.  
[http://opus.ipfw.edu/etcs\\_seniorproj\\_mctid/23](http://opus.ipfw.edu/etcs_seniorproj_mctid/23)

This Senior Design Project is brought to you for free and open access by the School of Engineering, Technology and Computer Science Design Projects at Opus: Research & Creativity at IPFW. It has been accepted for inclusion in Manufacturing & Construction Engineering Technology and Interior Design Senior Design Projects by an authorized administrator of Opus: Research & Creativity at IPFW. For more information, please contact [admin@lib.ipfw.edu](mailto:admin@lib.ipfw.edu).

FINAL REPORT ON THE CONSTRUCTION OF  
INSTRUMENTATION TO PERFORM IN-HOUSE TESTING  
OF TORQUE WRENCHES

prepared for

Professor Donald J. McAleece  
Department of Mechanical Engineering Technology  
Indiana-Purdue at Fort Wayne

prepared by

Erik A. Perry  
Mechanical Engineering Technology Student  
Indiana-Purdue at Fort Wayne

Date

April 5, 1988

This final report contains information in the areas of design, fabrication, testing, and cost of the in-house torque wrench testing instrumentation project.

## TABLE OF CONTENTS

INTRODUCTION. . . . .	1
DESIGN CRITERIA . . . . .	3
measurement, accurac, and temperature . . . . .	3
versatility . . . . .	4
portability . . . . .	5
MECHANICAL DESIGN AND COMPONENT SELECTION . . . . .	5
shaft . . . . .	6
shaft supports. . . . .	7
support platform. . . . .	8
strain instruments. . . . .	8
FABRICATION . . . . .	9
TESTING . . . . .	10
measurement range . . . . .	10
accuracy. . . . .	11
temperature . . . . .	12
versatility and portability . . . . .	13
COST ANALYSIS . . . . .	14
CONCLUSIONS AND RECOMMENDATIONS . . . . .	15
APPENDIX A. STRAIN GAGE INSTRUCTIONAL LITERATURE. . . . .	16
APPENDIX B. STRAIN GAGE INSTALLATION ACCESSORIES. . . . .	17
APPENDIX C. ENGLISH AND METRIC TORQUE COMPARISON. . . . .	18
APPENDIX D. 17-4 PH STAINLESS STEEL DATA. . . . .	19
APPENDIX E. SHAFT SIZE CALCULATION. . . . .	20
APPENDIX F. RECOMMENDED BEAM MATERIALS. . . . .	21
APPENDIX G. THE LRP-16 LINEAR-ROTARY BEARING. . . . .	22
APPENDIX H. THE THOMPSON PB-16-A PILLOW BLOCK . . . . .	23
APPENDIX I. THE P-3500 STRAIN INDICATOR . . . . .	24
APPENDIX J. THE P-3650 PEAK-READ INDICATOR. . . . .	25
APPENDIX K. ADDITIONAL STRAIN GAGE INSTRUMENTATION. . . . .	26
APPENDIX L. SAMPLE TEST RUN . . . . .	27
APPENDIX M. ITEMIZED COST OF 100-RUN LOT. . . . .	28
APPENDIX N. ITEMIZED COST OF PROTOTYPE. . . . .	29
APPENDIX O. SAMPLE TORQUE TRANSDUCER PRODUCT . . . . .	30
APPENDIX P. DETAIL AND ASSEMBLY DRAWINGS . . . . .	31
REFERENCES . . . . .	32

## INFORMATIVE ABSTRACT

The in-house torque wrench verification instrument gives the aviation mechanic confidence in the use of his torque wrenches. This instrument verifies if the torque wrench is in good operating condition thereby ensuring proper fastener installation. The instrument consists of three major parts: (1) transducer; (2) strain indicator; and (3) peak-read instrument. The transducer utilizes resistance-type strain gages to convert the mechanical input(torque) into an electrical output(millivolts). The strain indicator converts this electrical output into a strain reading. The peak-read instrument captures the maximum strain when the torque wrench reaches its preset value. The mechanic inserts the maximum strain in an equation and obtains a mathematical torque value applied by the wrench. This mathematical value is compared with the value indicated on the torque wrench. This comparison determines the operating condition of the torque wrench.